

FOVDAR: Sinusoidal Forced Vibration Test Controller

FOVDAR is a PC based hardware and software system that drives eccentric mass vibrators (EMV's) while analyzing feedback signals for amplitude and phase shift. FOVDAR-H is used for similar control of servo-hydraulic actuator vibrators. Using FOVDAR and vibrators allows the determination of modal properties of full sized civil and mechanical structures in the field. It uses a non-contact magnetic pulser to keep track of frequency and phase. All sensor input signals are analyzed with respect to that pulse. A user defined profile table defines the frequency sweep characteristics. There are numerous automatic and user controlled options for optimizing the sweep to increase data accuracy and reduce test time. All analogue input and output signals are generated or read with National Instruments (NI) hardware.

Features

- Output of analogue signals to drive one or multiple VFD motors/vibrators.
- 16 standard (expandable to 128) freely assignable analogue input channels to read a wide variety of sensors.
- Real time monitoring and display of force, eccentricity, and frequency.
- Real time data display of in phase component, out of phase component, amplitude modulus, and phase.
- Real time data analysis tools such as correlation verification allowing to see confidence levels of recorded data sets.
- Options for conversion from acceleration to velocity and displacement and normalization to force.
- Data set review feature without the need of running the EMV.
- Export data to MS Excel and as ASCII files to allow easy import into other programs.

The screenshot displays the FOVDAR 2.0 software interface with the following sections:

- Test Setup:** Includes tabs for Pulse Singal, Plot Data, and IP Channel Setup. A note indicates "FOVDAR 2.0 uses SI - Units [Deg, N, kg, m, s]".
- Profile Table:** A table with columns for Start Freq, End freq, Step Freq, Phase [Deg], and Step Type. A note above it says "Linear=0, Log >0 in %".
- Limits:**
 - Max Force: 1000 N
 - Max Freq: 1.48 Hz
 - Action Type: Warning Msg
- Pulser Setup:**
 - Pulse Ch: 1
 - Lead Angle: 0 deg CW
- Eccentricity:** 11.6193 kg m
- Sweep Stabilization Criteria:**
 - Max. Change: 50 %
 - Error: 5 %
 - Stab. Coef: 0.3
 - Stabilization: 3 Cycles
 - Average: 10 Cycles
 - Corr. Limit: 0.95
- Set VDV:** 1 V, Find Cal
- Set Freq:** 1 Hz, Set Freq
- Test Info:** A text area for test information.
- Loaded Profile:** A text field for the profile name.
- Base Path:** A text field for the base path.
- Base Filename:** A text field for the base filename.
- Footer Note:** "Saves all selected time histories for all set frequencies as per Profile Table"
- Right Panel:**
 - Dev# 10, DAQ-Type PCI-6229
 - Dev ID Dev1, # of AIs 32
 - Rate/Ch 7812, # of AOs 4
 - # of DI0s 48
 - VFD-Cal: 0.3543 V/Hz
 - VFD-Drive: 0 V
 - Force: 0 N
 - Table Entry #: 1
 - Start Freq: 10 Hz
 - End Freq: 3 Hz
 - Step Freq: 1 Hz
 - Set Freq: 0 Hz
 - Meas. Freq: 0 Hz
 - Running: [Green Circle]
 - Loop Status: [Green Circle]
 - Force Exceeded: [Green Circle]
 - Freq. Exceeded: [Green Circle]
 - Buttons: Save Graphs, Load ID, Save ID, Start Test, Reset, Exit, E-stop, Skip Freq, End Test

